

No Company Participated In this Tender - A

Technical evaluation report for equipment in TENDER-B

Key Parameters (Required)		M/S Alpha Technology Services	M/S Rastek Technologies
Item Type	6 1/2-Digit Bench Digital Multimeter	✓	✓
Model & OEM	6500, Keithley-Tektronix	✓	GDM-8261A, GW-Instek ✓
Measurement Functions	15 Measurement Function;	✓	11 Measurement Function; ✗
	• DCV: 100mV-1010V	✓	• DCV: 0.1μV-1000V
	• ACV: 100mV-750VRMS	✓	• ACV: 0.1μV-750VRMS
	• DCI: 10pA-10A ✓	✓	• DCI: 100pA-10A ✗
	• ACI: 100pA-10A	✓	• ACI: 1nA-10A
	• Resistance 2-wire: 10μΩ-120MΩ ✓	✓	• Resistance 2-wire: 100μΩ-100MΩ ✗
	• Resistance 4-Wire: 1μΩ-120MΩ ✓	✓	• Resistance 4-Wire: 100μΩ-100MΩ ✗
	• Capacitance: 0.1pF-120μF ✓	✓	• Capacitance: 0.1pF-120μF ✓
	• Period: 3.3μs-333ms ✓	✓	• Period: Range Not Mentioned in data sheet ✗
	• Frequency: 3Hz-300kHz ✓	✓	• Frequency: 3Hz-300kHz ✓
Display	• Diode: 10μV-12V	✓	• Thermocouple: -200C-1870C ✓
	• Digitize Voltage: 10μV-1010V	✓	• 2,3,4 Wire RTD: -200C-600C ✗
	• Digitize Current: 10nA-10A	✓	
	• Thermocouple: -200C-1820C ✓	✓	
	• Thermistor: -80C-150C	✓	
	• 2,3,4 Wire RTD: -200C-850C ✓	✓	
	• 5-inch multi-touch capacitive touchscreen with graphical display ✓	✓	• Florescent Display (VFD)
Data Transfer Rate Through USB	• 441000/s ✓	✓	• 24000/s ✗
Communication Interface:	• USB, GPIB, TSP-Link, and RS-232 ✓	✓	• USB, RS-232, Digital I/O ✗ • LAN, GPIB Optional
Internal Storage	• store up to 7 million readings ✓	✓	• No Internal Memory ✗
Software	• Built in internal Data Storage. Simply copy data from internal memory through USB stick to any PC. No external Software require • External PC link Software is available, free download	✓	• No built-in software. PC is required to interface meter to PC for instrument control ✗

Remarks:

- Equipment offered by M/S Rasiek Technologies, Model GDM-8261A is lower in specs as compared to Model 6500 as well as compared to specs as mentioned in tender.
- Equipment offer by M/S Alpha Technology Services fulfills the requirements of tender specs.

Decision:

Model: GDM-8261A is technically **REJECTED** due to lower specs.

Model: 6500 is technically **ACCEPTED** as it is according to tender demand and specs.


Dr. Hafizullah
Department of Physics
The Islamia University of Bahawalpur


Dr. Muhammad Azhar Khan
Department of Physics
The Islamia University of Bahawalpur

Dr. Muhammad Azhar Khan


Dr. Muhammad Azhar Khan
PI/Pakistan Science Foundation
Assistant Professor
Department of Physics
The Islamia University of Bahawalpur


Tender C

Technical assessment report for the Equipments at the Department of Physics

Sr. No.	Items	Quantity	Supplier	Specifications	Remarks
1	Digital Storage Oscilloscope	3	m/s Alpha Technology Services	Bandwidth = 70 MHz Channels = 2 Sample rate on each channel = 1.0 GS/s Record length = 2.5k points at all-time bases Model: TBS 1072B-EDU	1. The items offered by m/s Alpha Technology Services are according to the requirements. 2. The specifications are acceptable and can be procured for Electronics laboratory as soon as possible for fulfillment of experimentations.
2	Programmable 3-Ch DC Power Supply	2	m/s Alpha Technology Services	Maximum Voltage = 30V Maximum Current = 3A Maximum Power = 195W Model: 2230-30-3	1. The items offered by m/s Alpha Technology Services are according to the requirements. 2. The specifications are acceptable and can be procured for Electronics laboratory as soon as possible for fulfillment of experimentations.
3	5.5 Digit DMM	2	m/s Alpha Technology Services	Connectivity= with USB (TMC) / GPIB 220 Version Model: 2110-220-GPIB	1. The items offered by m/s Alpha Technology Services are according to the requirements. 2. The specifications are acceptable and can be procured for Electronics laboratory as soon as possible for fulfillment of experimentations.
6	BNC Cables	6	m/s RASTEK Technologies	Model: GTL-210	1. The items offered m/s RASTEK Technologies are according to the requirements. 2. The specifications are acceptable and can be procured for Electronics laboratory as soon as possible for fulfillment of experimentations.
8	Test Lead	20	m/s RASTEK Technologies	Model: GTL-103	1. The items offered m/s RASTEK Technologies are according to the requirements. 2. The specifications are acceptable and can be procured for Electronics laboratory as soon as possible for fulfillment of experimentations.
10	Digital Multimeter (Handheld)	5	m/s RASTEK Technologies	4000 counts Auto/Manual ranging Capacitance, Frequency measurement hFE test/EF Function Data hold and relative mode Auto power off With all standards accessories	1. The items offered m/s RASTEK Technologies are according to the requirements. 2. The specifications are acceptable and can be procured for Electronics laboratory as soon as possible for fulfillment of experimentations.


Dr. Hafeez Ullah Janjua,
 Assistant Professor.
 (Member)


Dr. Faris Iqbal,
 Assistant Professor.
 (Member)


Prof. Dr. Saeed Ahmad Buzdar,
 (Chairman).



786

The Islamia University of Bahawalpur
University College of Agriculture & Environmental Sciences
Department of Plant Pathology
E-mail: agriculture@iub.edu.pk, Ph: 062-9255531, Fax: 062-9250232

NO. 091/448ES

Date = 02/01/2020.

Subject: Evaluation of Quotations for tender-D/2019 Lab equipment's for the Department of Plant Pathology (Opened 27-12-2019).

It is stated that a tender was floated by The Islamia University of Bahawalpur for the Procurement of laboratory equipment's for the Department of Plant Pathology UCA & ES (Tender D/2019). The SMPC of the University in its meeting held on 27/12/2019, opened the technical proposals. The technical committee of the Department in its meeting held 30-12-2019 evaluated the technical proposals. The recommendations of the committee are as under

Item Description	Supplier	Model	Status	Remarks
1. Laminar Air Flow Hood Digital microprocessor control. Internal Dimension (WxDxH) in mm: 1200 x 645 x 610 Pre-filter: Disposable, non-washable polyester filter. Filter efficiency: > 99.999% at 0.3um Airflow velocity: average of 0.3 to 0.5m/s. Noise level: <60dBA. Fluorescent lamp: 28W. UV Lamp: 30W with safety system. Motorized front sash: Tempered glass, no less than 5mm, anti-ultraviolet radiation. Main Body Material: Adopting cold roll steel plate with corrosion proof coating with base stand. Power: 220V, 50Hz EEC, USA, UK, Japan	1. Scientific Technical Corporation (PVT) LTD	HC-VI20 BD-115	Accepted	As per Specification
2. Incubator Digital microprocessor controlled. Natural air convection, chamber capacity: 112 Liters, stainless steel interior & exterior, solid door, temperature range: 5°C above ambient temperature...+100°C, temperature resolution every 0.1°C, LCD graphic display with illuminated touch buttons, automatic air flap control, six-segment temperature time profile, 2 stainless steel wire shelves, access port, test results memory, open door alarm, door lock, logging with graph, internal glass door. EEC, USA, UK, Japan	2. Pakistan Analytic Supply	HPD-12204 XB112	Rejected	Not as per Specification
	3. S.T. Enterprises	RTVL 1312 CLN 155 STD	Accepted	As per Specification

Tender D

The Islamia University of Bahawalpur
 University College of Agriculture & Environmental Sciences
 Department of Plant Pathology
 E-mail: agriculture@iub.edu.pk, Ph: 062-9255531, Fax: 062-9250232




4. Premier Scientific Technologies	BBS-H1300 EN 120	Rejected	Not as per Specification

The selected firms are requested to attend the meeting of SMPC to held in future for opening of Financial Proposals.

Note: Quoted Item not according to the required specification

Departmental Technical Committee

1. Prof. Dr. Moazzam Jamil
2. Dr. Muhammad Naveed Aslam
3. Dr. Muhammad Naïees

Convenor 
 Member 
 Member 

Tender- E



The Islamia University of Bahawalpur

University College of Agriculture & Environmental Sciences
Department of Plant Breeding & Genetics
E-mail: pbguca@iub.edu.pk, Ph: 062-9255539, Fax: 062-9250232

Subject: Technical Comparative Statement for the Purchase of Laboratory Equipment (27-12-19)

Name of Company	Name of Equipment	Remarks
Azhan Enterprises	Ginning Machine	Selected
Khan Gee Enterprises	Ginning Machine	Selected
Ayyan Enterprises	Ginning Machine	Selected

Dr. Mueen Alam Khan

Dr. Muhammad Naeem

Prof. Dr. Moazzam Jamil

Prof. Dr. Muhammad Iqbal





The Islamia University of Bahawalpur

Department of Zoology

Ref. No: 896/Zoology

Date: 08-01-2020

Technical comparative statement for Tender-F/2019 Lab Equipment's (Zoology Department)
List of technical Approved Bidders for the purchase of Laboratory Equipment's in the Zoology Department.
Approved equipment's are recommended by the technical Committee to be considered for financial Assessment

Tender Opening Date: 27-Dec-2019

Sr No	Name of Items	Suppliers	Model/Make	Status	Remarks
1	Hematology Analyzer/Blood Analyzer	Scientific Technical corporation Lhr			NQ
		Rays Technology Lhr			NQ
		Premier Scientific Technology Rwp			NQ
		Vision Scientific Corporation	AHA-3000VET Acculab, USA	Accepted	
2	UV/Vis Double Beam Spectrophotometer	Scientific Technical corporation Lhr	C-7200S Peak Instrument Inc, USA	Rejected due to. Wave length can't matched	Not as per Specification
		Rays Technology Lhr	Specord 200 Plus. Analitikajena, Germany	Rejected due to. Wave length can't matched	Not as per Specification
		Premier Scientific Technology Rwp.	UVD-2950 Labomed Inc, USA	Rejected	Not as per specification
		Vision Scientific Corporation	UV-1602 BMS, Canada	Accepted 8 cell Holder	
3	Multi Parameter	Scientific Technical corporation Lhr	S-630 Peak Instrument USA	Accepted	
		Rays Technology Lhr	Multi3620 IDS WTW, Germany	Rejected	Not as per specification
		Premier Scientific Technology Rwp			NQ
		Vision Scientific Corporation	HI2030 HANNA Instrument, USA	Accepted	
4	Digital Analytical Balance	Scientific Technical corporation Lhr	SAB-414i ADAM Equipment Co, UK	Accepted	
		Rays Technology Lhr	ATX-324 Shimadzu, Japan	Accepted	
		Premier Scientific Technology Rwp	B314Ai Optika, Italy	Accepted	
		Vision Scientific Corporation	BA2204B BioBase, Korea	Rejected	Not as per specification

Departmental Purchase Committee:

Member 1
Prof. Dr. Nuzhat Sial

Department of Zoology
The Islamia University of Bahawalpur

Member 2
Dr. Abdul Ghaffar

Associate Professor
Department of Zoology
The Islamia University of Bahawalpur

Member 3
Dr. Mushtaq Hussain Lashari

Assistant Professor
Department of Zoology
The Islamia University of Bahawalpur

Dr. Abdul Ghaffar
Project Manager
PARB Project No. 689
Department of Zoology
The Islamia University of Bahawalpur



Scanned with
CamScanner



Technical Statement for the Purchase of Glassware

Sr.#	Name of Chemicals	Wahid Scientific Store Lahore	Price Inclusive of G.S.T	M/S RAHAT Enterprises Bahawalpur	Price Inclusive of G.S.T	ILYAS Sons Scientific Co. Lahore	Price Inclusive of G.S.T	Usman Lab Traders Faisalabad	Price Inclusive of G.S.T
1	Beaker, 2 Liter	Joan Lab		China		China		Pyrex/China	
		Accepted		Accepted		Accepted		Accepted	
2	Beaker, 1000ml	Joan Lab		China		China		Pyrex/China	
		Accepted		Accepted		Accepted		Accepted	
3	Beakers,, 500ml	Joan Lab		China		China		Pyrex/China	
		Accepted		Accepted		Accepted		Accepted	
4	Beakers, 250ml	Joan Lab		China		China		Pyrex/China	
		Accepted		Accepted		Accepted		Accepted	
5	Beakers, 100ml	Joan Lab		China		China		Pyrex/China	
		Accepted		Accepted		Accepted		Accepted	
6	Falcon Tubes 10ml 50 Pcs in 1 Pkt	Imported		China		China		Bio Basic Inc.	
		Accepted		Accepted		Accepted		Accepted	
7	Falcon Tubes 10ml 50 Pcs in 1 Pkt	Imported		China		China		Bio Basic Inc	
		Accepted		Accepted		Accepted		Accepted	
8	Measuring Cylinder 100ml	Joan Lab		China		China		China	
		Accepted		Accepted		Accepted		Accepted	
9	Beaker 5 Liter	Joan Lab		China		China		China	
		Accepted		Accepted		Accepted		Accepted	
10	Measuring Cylinder 10ml	Joan Lab		China		China		China	
		Accepted		Accepted		Accepted		Accepted	
11	Conical Flask, 250ml	Joan Lab		China		China		China	
		Accepted		Accepted		Accepted		Accepted	
12	Petri Dishes 4"	China		China		China		China/Normax/EU	
		Accepted		Accepted		Accepted		Accepted	
13	Eppendroph Tubes (1.5ml) 500 pcs in 1 pkt.	Imported		China		China		China	
		Accepted		Accepted		Accepted		Accepted	
14	Glass Slides	China		China		China		China	
		Accepted		Accepted		Accepted		Accepted	


1963
06/10/2020


ASSISTANT PROFESSOR
Department of Pharmacy
Government College University
Faisalabad


ASSISTANT PROFESSOR
Department of Pharmacy
Government College University
Faisalabad

ASSISTANT PROFESSOR
Department of Pharmacy
Government College University
Faisalabad

32	Spatula	Pak	China	China	China / Pak
33	Receiving Adapter 19/26	Rejected Pyrex	Accepted China	Accepted China	Accepted Pyrex / ISO - Lab
34	Round Button Flask 29/32 1liter	Accepted China	Accepted China	Accepted Pyrex	Accepted Pyrex / ISO - Lab
35	Receiving Flask 19/26 500ml	Accepted China	Accepted China	Accepted China	Accepted Pyrex / ISO - Lab
36	TLC Tank	Accepted China	Accepted China	Accepted China	Accepted China Germany
37	Reagent Bottle 250ml	Accepted China	Accepted China	Accepted China	Accepted China
38	Capillary Tubes	Accepted China	Accepted China	Accepted Pak	Accepted Germany
39	Test Tube Holder	Accepted France	Accepted China	Rejected China	Accepted China / Pak
40	Whatman Filter Paper (40,41 & 42)	Accepted China	Accepted China	Accepted Diamond	Accepted Whatman / UK
41	Aluminum Foil	Accepted Russia	Accepted China	Accepted Pyrex	Accepted Diamond / USA
42	Quartz UV Cell 10mm	Accepted France	Rejected China	Accepted Pyrex	Accepted ISO - Lab / Germany
43	Cellulose Membrane Filter (0.45 Micro) 100pcs in 1Pkt	China	China	China	Accepted Membrane Solution
44	China Dish (6 inches)	Rejected Russia	Rejected China	Rejected China	Accepted China
45	Centrifuge Tubes (Glass)	Accepted France	Accepted China	Accepted China	Accepted China


 CHAIRMAN
 Department of Pharmacy
 The American University of Beirut



 LECTURER
 Department of Pharmacy
 The American University of Beirut



 ASSOCIATE PROFESSOR
 Department of Pharmacy
 The American University of Beirut




Scanned with

15	Glass Covers	China Accepted		China Accepted		China Accepted		China Accepted
16	Pestle & Mortar (5 Inch)	China Accepted		China Accepted		China Accepted		China Accepted
17	Funnel (4 Inch)	China Accepted		China Accepted		China Accepted		China Accepted
18	Test Tube (20ml)	China Accepted		China Accepted		China Accepted		China Accepted
19	Test Tube Stand Steel Pak	China Accepted		China Accepted		China Accepted		China Accepted
20	Pipette 1ml	Rejected		China Accepted		China Accepted		China Accepted
21	Pipette 5ml	Germany Accepted		China Accepted		China Accepted		China Accepted
22	Pipette 10 ml	Germany Accepted		China Accepted		China Accepted		China Accepted
23	Still Head Adapter 19/26 + Stopper 19/26	Pyrex Accepted		China Accepted		China Accepted		China Accepted
24	Round Button Flask 19/26 500ml	Joan Lab Accepted		China Accepted		China Accepted		China Accepted
25	Round Button Flask 19/26 250ml	Joan Lab Accepted		China Accepted		China Accepted		China Accepted
26	Round Button Flask 19/26 100ml	Joan Lab Accepted		China Accepted		China Accepted		China Accepted
27	Magnetic Stirrer bar (1inch & ½ inch)	Germany Accepted		China Accepted		China Accepted		China Accepted
28	Iron Stand (Complete Set)	Pak Rejected		China Accepted		China Accepted		China Accepted
29	Condenser 19/26	China Accepted		China Accepted		China Accepted		China Accepted
30	Burner	Pak Rejected		China Accepted		China Accepted		China Accepted
31	Tripod Stand	Pak Rejected		China Accepted		China Accepted		China Accepted


 ASSOCIATE PROFESSOR
 Department of Pharmacy
 Qu. 3, Sector 3, Gurgaon, Haryana


 LECTURER
 Department of Pharmacy
 Qu. 3, Sector 3, Gurgaon, Haryana


 CHAIRMAN
 Department of Pharmacy
 Qu. 3, Sector 3, Gurgaon, Haryana

Technical Statement for the Purchase of Chemicals

Sr.#	Name of Chemicals	Wahid Scientific Store Lahore	Price Inclusive of G.S.T	Musaji Adam & Sons Multan	Price Inclusive of G.S.T	M/S RAHAT Enterprises Bahawalpur	Price Inclusive of G.S.T	ILYAS Sons Scientific Co. Lahore	Price Inclusive of G.S.T	Usman Lab Traders Faisalabad	Price Inclusive of G.S.T
1	Acetonitrile (HPLC) Grade	Sigma/ Aldrich Accepted		Daejun, Korea Rejected		China/Korea Rejected		Sigma Accepted		Samchun/Korea Rejected	
2	Methanol (HPLC) Grade	Sigma/ Aldrich Accepted		Daejun, Korea Rejected		China/Korea Rejected		Sigma Accepted		Samchun/Korea Rejected	
3	Chloroform (Analytical Grade)	Sigma/ Aldrich Accepted		Daejun, Korea Rejected		China/Korea Rejected		Sigma Accepted		MERCK Accepted	
4	Falcon Tube (15ml)	Imported Accepted		Axygen - Corning Accepted		China/Korea Accepted		China Accepted		Bio Basic Inc. Accepted	
5	Falcon Tube (50ml)	Imported Accepted		Axygen - Corning Accepted		China/Korea Accepted		China Accepted		Bio Basic Inc. Accepted	
6	Appendorph (Tube 1.5ml)	Imported Accepted		Kartell Accepted		China/Korea Accepted		China Accepted		China Accepted	
7	Magnetic Stirrer bar ½ inch Magnetic Stirrer bar 3 1 inch	Germany Accepted				China/Korea Accepted		China Accepted		Germany Accepted	
8	Adjustable Sucker	DLAB USA Accepted		Kartell Accepted		China/Korea Accepted		China Accepted		China Accepted	
9	N. Butanol (Commercial)	France Accepted		Daejun, Korea Accepted		China/Korea Accepted		China Accepted		BDH Accepted	
10	Methanol (Commercial)	Saudia Accepted		Daejun, Korea Accepted		China/Korea Accepted		China Accepted		Accepted GULAF	
11	Ethanol (Commercial)	Ravi Rayan Accepted				China/Korea Accepted		China Accepted		Accepted Lab IMP Grade	
12	N. Butanol (Commercial)	France Accepted				China/Korea Accepted		China Accepted		Accepted Korea	
13	N. Hexane (Commercial)	China Accepted		Daejun, Korea Accepted		China/Korea Accepted		China Accepted		Accepted Saudi Arabia	
14	Chloroform (Commercial)	Russia Accepted		Daejun, Korea Accepted		China/Korea Accepted		China Accepted		Accepted Russia	

ASSOCIATE PROFESSOR
Department of Pharmacy
The University of Balochistan

LECTURER

Department of Pharmacy
The University of Balochistan

CHAPMAN
Department of Pharmacy
The University of Balochistan



Scanned with
CamScanner



